

## SoundEar®3

## Measure, monitor and manage the noise with SE3-300

## SoundEar®3 - 300 Specifications

LAF; LAS; LC; Laeq 1/2h, La Q,1 dB for all p Measuring Ranges: RMS: Total 30 Deviation: +/- 0,5 dB Frequency Range: 20Hz- 20 kHz Frequency Weightings: A- weighting Time Weighting: Slow(1S) & Fa Dynamic Range: 90 dB and per Light setting: full configura including nigh 2 x outputs : 0-10 V or 4-20 2 xUSB outputs: Micro USB ( p ( Log, configu Display setting: LAS max, Leo

Power Supply: Current Consumption: Internal memory:

Parameters:

Measures 3 measurements simultaneously LAF; LAS; LCpeak; Laeg, 1s, Laeg 1/4 h, Laeq 1/2h, Laeq 1 h. 0,1 dB for all parameters RMS: Total 30 - 120 dB 20Hz- 20 kHz A-weighting (RMS), C-weighting (Peak) Slow(1S) & Fast (125 ms) 90 dB and peak detection full configurability through Soundear software including nightsetting. 0-10 V or 4-20 mA Micro USB (power & PC), USB OTG (Log, configuration) LAS max, Leq (A)15, Alarm settings, Temperature & Clock 5VDC (Micro USB) / 24VDC (Screw terminal) max 2,5 W 16 MB(128 Mbit) (5-90 days log time,

depending on log settings )



The mini display on SoundEar®3 – 300 and 310 is operated manually via the tuch panel on the front of the cabinet.



Eksternal microphone.



USB key for export of data.



SoundEar®3 software



Real Time Clock: Microphone: Measurement 300 - 310:

Hi-precision type with battery backup (CR2032) 20 Hz- 20 kHz

Length 256 mm, Width: 205 mm, Height: 45 mm, Weight: 1,5 kg

**Standards:** IEC61672-2-2002, Type 2, ANSI 51,4 Type 260601-1: Medical electrical equipment- Part 1: general requirements for basic safety and essential performance. 60601-1-2\_ Medical equipment – Part 1.2: General requirement for Basic safety and essential performance.

**Connectivity accessories:** GSM module, 4 G module for Cloud solution